

Daftar Pustaka

- [1] Ub521 datasheet. https://www.homecontrols.com/homecontrols/products/pdfs/AN-AnD/ANUB521_datasheet.pdf, 2008. Online; Diakses pada 20 Juni 2018.
- [2] D. G. Altman and J. M. Bland. Standard deviations and standard errors. *Bmj*, 331(7521):903, 2005.
- [3] K. Chan, K. Hung, and Y. Zhang. Noninvasive and cuffless measurements of blood pressure for telemedicine. In *Engineering in Medicine and Biology Society, 2001. Proceedings of the 23rd Annual International Conference of the IEEE*, volume 4, pages 3592–3593. IEEE, 2001.
- [4] K. Eguchi, S. Kuruvilla, G. Ogedegbe, W. Gerin, J. E. Schwartz, and T. G. Pickering. What is the optimal interval between successive home blood pressure readings using an automated oscillometric device? *Journal of hypertension*, 27(6):1172, 2009.
- [5] S. S. Franklin. The importance of diastolic blood pressure in predicting cardiovascular risk. *Journal of the American Society of Hypertension*, 1(1):82–93, 2007.
- [6] H. Gesche, D. Grosskurth, G. Küchler, and A. Patzak. Continuous blood pressure measurement by using the pulse transit time: comparison to a cuff-based method. *European journal of applied physiology*, 112(1):309–315, 2012.
- [7] S. Goli and T. Jayanthi. Cuff less continuous non-invasive blood pressure measurement using pulse transit time measurement. *Int. J. Recent Dev. Eng. Technol*, 2:86–91, 2014.
- [8] A. Hennig and A. Patzak. Continuous blood pressure measurement using pulse transit time. *Somnologie-Schlafforschung und Schlafmedizin*, 17(2):104–110, 2013.
- [9] G. Jernstedt and J. P. Newcomer. Blood pressure and pulse wave velocity measurement for operant conditioning of autonomic responding. *Behavior Research Methods & Instrumentation*, 6(4):393–397, 1974.
- [10] Z. Jin-ling, L. Yue, and W. Jia-bao. Design of electronic blood pressure monitoring system based on mobile telemedicine system. In *Complex Medical Engineering (CME), 2013 ICME International Conference on*, pages 145–149. IEEE, 2013.
- [11] Ł. Mokros, W. Kuczyński, Ł. Franczak, and P. Białasiewicz. Morning diastolic blood pressure may be independently associated with severity of obstructive sleep apnea in non-hypertensive patients: a cross-sectional study. *Journal of clinical sleep medicine: JCSM: official publication of the American Academy of Sleep Medicine*, 13(7):905, 2017.
- [12] E. Park, B. Cho, S. Park, J. Lee, J. Lee, I. Kim, and S. I. Kim. Continuous measurement of systolic blood pressure using the ptt and other parameters. In *Engineering in Medicine and Biology Society, 2005. IEEE-EMBS 2005. 27th Annual International Conference of the*, pages 3555–3558. IEEE, 2006.
- [13] L. Peter, N. Noury, and M. Cerny. A review of methods for non-invasive and continuous blood pressure monitoring: Pulse transit time method is promising? *Irbm*, 35(5):271–282, 2014.
- [14] C. Poon and Y. Zhang. Cuff-less and noninvasive measurements of arterial blood pressure by pulse transit time. In *Engineering in Medicine and Biology Society, 2005. IEEE-EMBS 2005. 27th Annual International Conference of the*, pages 5877–5880. IEEE, 2006.
- [15] H. Rasyid and R. St Rabiul Zatalia. Hubungan obstructive sleep apnea dengan hipertensi.
- [16] H. Sheng, M. Schwarz, and J. Boercsoek. A wireless wearable body sensor network for continuous noninvasive blood pressure monitoring using multiple parameters. *measurement*, 6:17, 2011.
- [17] X. Teng and Y. Zhang. Continuous and noninvasive estimation of arterial blood pressure using a photoplethysmographic approach. In *Engineering in Medicine and Biology Society, 2003. Proceedings of the 25th Annual International Conference of the IEEE*, volume 4, pages 3153–3156. IEEE, 2003.
- [18] S.-y. Ye, G.-R. Kim, D.-K. Jung, S. Baik, and G. Jeon. Estimation of systolic and diastolic pressure using the pulse transit time. *World Academy of Science, Engineering and Technology*, 67:726–731, 2010.