**ABSTRACT** 

Prolonged lying down can cause pressure ulcers, especially in bedridden patients with

limited mobility. Pressure ulcers are triggered by continuous pressure on specific points of the

body, with key factors including temperature and pressure while lying down. This capstone

design aims to develop a monitoring system for temperature and pressure to prevent pressure

ulcers in bedridden patients. Testing results indicate that the FSR402 pressure sensor has

accuracy issues due to its small surface area, requiring recalibration or replacement.

Meanwhile, temperature measurements show that the patient's body temperature remains

within the normal range (36–37.6°C) but tends to increase with prolonged lying down.

Therefore, it is recommended that patients be repositioned every two hours to prevent excessive

temperature rise and further complications.

Keywords: Bed Rest, Patients, Mobility, Temperature, Pressure.

xiv